



Shenzhen GCT Technical Service Co.,Ltd.

HOTLINE: 400-096-8018

WWW.GCT-LAB.COM

Certificate of Compliance

Certificate No. : GCT200478072E
Applicant : ShenZhen Changkun Technology CO.,LTD
Applicant Address : 3ffloor 4ffloor 5ffloor 7ffloor,B Building,NO.69,Zhenbi Road,Biling Community,Pingshan Street,Pingshan District,Shenzhen City,Guangdong,China
Manufacturer : ShenZhen Changkun Technology CO.,LTD
Manufacturer Address : 3ffloor 4ffloor 5ffloor 7ffloor,B Building,NO.69,Zhenbi Road,Biling Community,Pingshan Street,Pingshan District,Shenzhen City,Guangdong,China
Product : Pulse Oximeter
M/N : X1805
X1906,X1907,X1908,X1909
Trademark : N/A
Test Standard : EN 60601-1-2:2015
(CISPR 11:2015;IEC 61000-3-2:2014;IEC 61000-3-3:2013;
IEC 61000-4-2:2008;IEC 61000-4-3:2006;IEC 61000-4-4:2012;
IEC 61000-4-5:2014;IEC 61000-4-6:2013;IEC 61000-4-8:2009;
IEC 61000-4-11:2004)

The EUT described above has been tested by us with the listed standards and found in compliance with the council MDD directive 93/42/EEC. This document has been issued on a voluntary basis and upon request of the manufacturer. It is possible to use CE marking to demonstrate the compliance with this MDD Directive. It is only valid in connection with the test report number: GCT200478072ER.

CE



Manager

Issue Date: Apr. 21, 2020