



Certificate of Conformity

Certificate Number: DL-20210924021C

Applicant: Shenzhen Deda Medical Technology Group Co., Ltd.

Building 39-1, Dayun Software Town, Longgang Avenue, No.8288, Henggang street, Longgang District, Shenzhen, Guangdong, 518100, P.R.China

Manufacturer: Deda Medical (Hunan) Co., Ltd.

Deda Industrial Park, Jianshe Avenue, Huaihua High-tech Zone, Hunan Province

Product: Homecare oxygen concentrator

Model No.: DE-2SW

DE-Q1, DE-Q1W, DE-1SW, DE-1S, DE-2S, DE-1LW, DE-1A, DE-1B, DE-2AW, DE-2A, KH21-T10, KH21-T12, DE-Y3AW, DE-Y3A, DE-Y5AW, DE-Y5A

Test Standard: EN 55014-1:2017/A11:2020

EN IEC 61000-3-2:2019+A1:2020, EN 61000-3-3:2013+A1:2019

EN 55014-2:2015

EN 61000-4-2:2009, EN IEC 61000-4-3:2020, EN 61000-4-4:2012,

EN 61000-4-5:2014+A1:2017, EN 61000-4-6:2014/AC:2015,

EN IEC 61000-4-11:2020

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. It is only valid in connection with the test report number: DL-20210924021E.



This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole product and relevant. Without the written approval, It is not permitted to use the test lab's logo.

Shenzhen DL Testing Technology Co., Ltd.

101-201, Building C, Shuanghuan, No.8, Baoqing Road, Baolong Industrial Zone, Baolong Street, Longgang District, Shenzhen, Guangdong, China

Web: www.dl-cert.com E-mail: Service@dl-cert.com Tel: 400-688-3552

